

University of Dundee

## Implant Design Affects Walking and Stair Navigation after Total Knee Arthroplasty

Komaris, Dimitrios Sokratis; Govind, Cheral; Murphy, Andrew James; Clarke, Jon; Ewen, Alistair; Leonard, Hollie

*Published in:*  
Journal of Orthopaedic Surgery and Research

*DOI:*  
[10.1186/s13018-021-02311-x](https://doi.org/10.1186/s13018-021-02311-x)

*Publication date:*  
2021

*Licence:*  
CC BY

*Document Version*  
Publisher's PDF, also known as Version of record

[Link to publication in Discovery Research Portal](#)

*Citation for published version (APA):*  
Komaris, D. S., Govind, C., Murphy, A. J., Clarke, J., Ewen, A., Leonard, H., & Riches, P. (2021). Implant Design Affects Walking and Stair Navigation after Total Knee Arthroplasty: a double-blinded randomised controlled trial. *Journal of Orthopaedic Surgery and Research*, 16, [177]. <https://doi.org/10.1186/s13018-021-02311-x>

### General rights

Copyright and moral rights for the publications made accessible in Discovery Research Portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from Discovery Research Portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain.
- You may freely distribute the URL identifying the publication in the public portal.

### Take down policy


If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

RESEARCH ARTICLE

Open Access



# Implant design affects walking and stair navigation after total knee arthroplasty: a double-blinded randomised controlled trial

Dimitrios-Sokratis Komaris<sup>1,2\*</sup> , Cheral Govind<sup>2</sup>, Andrew James Murphy<sup>2,3</sup>, Jon Clarke<sup>4</sup>, Alistair Ewen<sup>4</sup>, Hollie Leonard<sup>4</sup> and Philip Riches<sup>2</sup>

## Abstract

**Background:** Dissimilar total knee arthroplasty implant designs offer different functional characteristics. This is the first work in the literature to fully assess the Columbus ultra-congruent mobile (UCR) system with a rotating platform.

**Methods:** This is a double-blinded randomised controlled trial, comparing the functional performance of the low congruent fixed (CR DD), ultra-congruent fixed (UC) and UCR Columbus Total Knee Systems. The pre-operative and post-operative functional performance of twenty-four osteoarthritic patients was evaluated against nine control participants when carrying out everyday tasks. Spatiotemporal, kinematic and kinetic gait parameters in walking and stair navigation were extracted by means of motion capture.

**Results:** The UC implant provided better post-operative function, closely followed by the UCR design. However, both the UC and UCR groups exhibited restricted post-operative sagittal RoM (walking,  $52.1 \pm 4.4^\circ$  and  $53.2 \pm 6.6^\circ$ , respectively), whilst patients receiving a UCR implant did not show an improvement in their tibiofemoral axial rotation despite the bearing's mobile design (walking, CR DD  $13.2 \pm 4.6^\circ$ , UC  $15.3 \pm 6.7^\circ$ , UCR  $13.5 \pm 5.4^\circ$ ). Patients with a CR DD fixed bearing showed a statistically significant post-operative improvement in their sagittal RoM when walking ( $56.8 \pm 4.6^\circ$ ).

**Conclusion:** It was concluded that both ultra-congruent designs in this study, the UC and UCR bearings, showed comparable functional performance and improvement after TKA surgery. The CR DD group showed the most prominent improvement in the sagittal RoM during walking.

**Trial registration:** The study is registered under the clinical trial registration number: [NCT02422251](https://clinicaltrials.gov/ct2/show/study/NCT02422251). Registered on April 21, 2015.

**Keywords:** Knee prosthesis, Fixed bearing, Mobile bearing, Implant congruency, Range of motion, Motion analysis

\* Correspondence: [sokratis.komarís@tyndall.ie](mailto:sokratis.komarís@tyndall.ie)

<sup>1</sup>Tyndall National Institute, University College Cork, Lee Maltings Complex Dyke Parade, Cork T12 R5CP, Ireland

<sup>2</sup>Department of Biomedical Engineering, University of Strathclyde, Glasgow, Scotland

Full list of author information is available at the end of the article



© The Author(s). 2021 **Open Access** This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>. The Creative Commons Public Domain Dedication waiver (<http://creativecommons.org/publicdomain/zero/1.0/>) applies to the data made available in this article, unless otherwise stated in a credit line to the data.

## Background

Advancements in the design of total knee arthroplasty (TKA) prostheses have led to the commercialisation of more than 150 types of knee implants [1] with potentially different functions. For example, in fixed-bearing designs, the polyethylene sheet is fixed upon the underlying tibial component, whilst in mobile bearings the insert can rotate short distances inside the metal tibial tray; bearing congruency also changes the level of conformity between the femoral section and the bearing surface, which in theory affects mobility, contact forces, and polyethylene wear. The possible advantages of mobile over fixed bearings (e.g., [2]), posterior cruciate ligament (PCL) retention versus substitution (e.g., [3]), and the clinical outcomes of different bearing congruencies (e.g., [4]), component fixation methods (e.g., [5]) and patellar resurfacing techniques (e.g., [6]) have all been extensively investigated. In the present study, three types of Columbus® total knee prostheses (B. Braun Aesculap, Tuttlingen, Germany) with different platform designs, degrees of congruency and PCL management are compared.

Variations of the Columbus knee have been previously investigated, either between different bearing designs or in comparison to other commercially available implants

with no statistical differences found between mobile and fixed bearings, and with similar Oxford Knee Scores and passive sagittal range of motion (RoM) between all investigated commercially accessible prostheses (Table 1). Additionally, studies investigating other commercially available knee implants have had similar findings. For example, Urwin et al. [18] reported no significant differences between fixed (Sigma® Fixed Bearing Knee System, De Puy International) and mobile bearings (Sigma® Rotating Platform Knee System, DePuy International) in spatiotemporal, kinematic and kinetic measurements (stride length and time, gait velocity, flexion angles during walking, knee RoM and maximum knee adduction moment) at 9 months post-operatively. Additionally, TKA patients in the same study walked with greater minimum knee flexion and reduced knee adduction moment when compared to the control group.

However, commonly used clinical scores, such as knee RoM and patient-reported outcome measures, are insensitive and can be inadequate due to the presence of floor and ceiling effects [19]. Quantitative biomechanical analyses of TKA patients during activities of daily living, including more functionally challenging activities such as stair ascent and descent, may be more appropriate and statistically powered to evidence differences between

**Table 1** Summary of papers with Columbus knee implants

Study	Prostheses	Type of bearing	No of knees	Knee passive flexion RoM (°)		Significant differences found
				Pre-operative	Post-operative	
Lampe et al. [7]	Columbus CR	Fixed	52	111 ± 15	113 ± 13	No <sup>a</sup>
	Columbus RP	Mobile	48	109 ± 12	115 ± 11	
Goebel and Schultz [8]	Columbus Knee	Fixed	109	-	-	Yes <sup>c</sup>
	NexGen Full Flex	Fixed	22	-	-	
Jung et al. [9]	Columbus PS	Fixed	197	128.4 ± 16.2	131.8 ± 10.7	No <sup>c</sup>
	Scorpio PS	Fixed	187	124.5 ± 19.9	130.2 ± 14.2	
Hakki et al. [10]	Columbus DD/UC	Fixed	79	94.7	110.4	-
Marques, Daniel [11]	Columbus CR	Fixed	45	110.6 ± 15.5	114.3 ± 9.3	No <sup>a</sup>
	Columbus RP	Mobile	42	109.4 ± 12.7	117.7 ± 10.9	
Luzo et al. [12]	Columbus PS	Fixed	196	-	-	Yes <sup>b</sup>
Kim et al. [13]	Columbus UC	Fixed	73	115.5 ± 10.3	125.6 ± 9.1	No <sup>a</sup>
	E-motion UC	Mobile	73	114.1 ± 9.6	123.7 ± 9.7	
Lutzner et al. [14]	Columbus UC	Fixed	63	102.9 ± 14.6	112.2 ± 11.8	No <sup>c</sup>
	Columbus PS	Fixed	64	102.1 ± 13.1	115.1 ± 13.0	
Yoon and Yang [15]	Columbus UC	Fixed	233	125.0 ± 13	-	-
Yoon and Yang [16]	Columbus UC	Fixed	105	125 ± 13	127 ± 8	No <sup>a</sup>
	E-motion FP	Mobile	95	129 ± 10	120 ± 5	
Fuchs et al. [17]	Columbus CRDD	Fixed	187	106.3 ± 20.2	114.0 ± 12.1	Yes <sup>b</sup>

CR cruciate retaining, DD deep dish, RP rotating platform, PS posterior stabilised, UC ultra-congruent, FP floating platform

<sup>a</sup>Between fixed and mobile bearings

<sup>b</sup>Between pre-operative and post-operative assessment

<sup>c</sup>Between different commercially available knee designs

prostheses. To further explore whether implant design paradigm variations result in significant differences, this study evaluates the full biomechanical performance of three different bearing configurations of Columbus knee replacement implants and the age-matched natural knee; the hypothesis being that the mobile high congruent bearing will facilitate more natural movements during everyday tasks.

## Material and methods

### Recruitment

Following appropriate ethical approvals, volunteers were recruited to this randomised, controlled, double-blinded study ([ClinicalTrials.gov](https://clinicaltrials.gov/ct2/show/study/NCT02422251) identifier: NCT02422251). Volunteers were sought from patients scheduled for unilateral TKA at the Golden Jubilee National Hospital, Clydebank, Scotland, between August 2015 and June 2017. Patients were excluded if they had had a hip or knee replacement procedure in the previous 12 months, had previous ankle surgery or past neurologic history (e.g., stroke). Recruits gave written informed consent and were blindly randomised using sequentially numbered opaque sealed envelopes, to receive one of three designs: a low congruent fixed (CR DD, cruciate-retaining deep-dish), a high congruent fixed (UC, ultra-congruent) or a high congruent mobile (UCR, ultra-congruent rotating platform) knee bearing (Columbus Total Knee Systems, B. Braun, Melsungen, Germany). A nominated person, independent of the approach and consent of the patient, opened the envelopes and informed the hospital team of the randomisation. Both the patient and the research team responsible for the motion capture and data analyses were blinded to the implant allocation.

The high congruent bearings used in the study are posterior stabilised and require the posterior cruciate ligament (PCL) to be resected; the CR DD on the other hand has a cruciate retaining design. All tibial and femoral implant components were fixed with cement, and surgeries were performed using the OrthoPilot navigation system. If, intra-operatively, the allocated bearing was deemed inappropriate for a given patient, the operating surgeon could choose a more suitable implant, and the patient was excluded from the study. Operations were performed by three consultant orthopaedic surgeons who all used the three study implants in their routine clinical practice. Age-matched, asymptomatic, control volunteers were recruited from community groups and social clubs. Control participants were excluded if they had a previous lower limb joint replacement procedure, ankle surgery, or any musculoskeletal, neurological, or sensory deficit.

Seven hundred and forty-four patients were assessed for eligibility, of which forty-two patients were recruited in the study (Fig. 1), along with nine control participants.

Patients' biomechanical performance was assessed pre-operatively and 1 year post-operatively in levels walking, stair ascent, and stair descent. Control participants attended a single motion capture session only. Out of forty-two, eight patient participants failed to attend their pre-operative assessment at the designated time period of 1 week before their operation and were excluded from the study; another eight patients withdrew post-operatively for health reasons; finally, two patients did not receive a Columbus implant upon surgery. Therefore, a total of twenty-four patients were included in this analysis, of which, twelve, five, and seven received the CR DD, UC, and UCR implants, respectively.

### Demographics


Gender, age, and body mass index (BMI) were recorded for all groups, and affected knee side was noted for the patient groups (Table 2). Pre- and post-operative Oxford Knee Scores (OKS) were also logged by clinicians during hospital visits. A patient's OKS is generated from a twelve-item questionnaire, assessing the difficulty and pain levels during common activities of daily living; the best possible outcome is 12, and the score giving the worst is 60.

Age, BMI, and OKS between groups were compared using one-way ANOVA tests with a Bonferroni correction. Patients showed significantly higher BMI values than the control group ( $P < 0.01$ ). No statistically significant differences in age and OKS were noted between all groups. *T*-tests showed no change in BMI post-operatively; OKS was significantly improved after surgery for all implant groups ( $P < 0.01$ ). Patient satisfaction was also assessed 1-year after TKA using a five-point Likert scale (1, very dissatisfied; 2, dissatisfied; 3, unsure; 4, satisfied; 5, very satisfied), with twenty-two and two patients reporting to be very satisfied and satisfied, respectively.

### Instrumentation

Motion capture was carried out with twelve Vicon (Oxford Metrics, Oxford, UK) T-series cameras and four Kistler (Winterthur, CH) force platforms, sampling at 100 and 1000 Hz, respectively. Prior to each recording, a calibration weight of 200 N was placed on top of each force plate, and the error in the recorded value was logged. Following the calibration of the motion capture system, the highest camera error (mm) was also noted. Both errors were low and consistent across all recordings (force plates  $1.9 \pm 0.9\%$ , cameras  $0.19 \pm 0.009$  mm).

Male participants wore Lycra shorts and comfortable footwear; female participants additionally wore Lycra t-shirts. Following anthropometric measures, thirty-five 14-mm retroreflective markers were fixed to anatomical body locations as per the full-body Plug-in Gait



biomechanical model. A knee alignment device (KAD) was used during the static calibration of the participants.

### Recordings

Initially, participants walked for 10 m at a comfortable walking speed. No other instructions or information about the existence of force platforms was given. Both starting and finishing positions were clearly marked on the floor with a coloured tape. A minimum of six walking trials were captured per participant, with extra being

recorded if the number of clean force plate strikes was lower than five.


For the stair ascent and descent tasks, a four-step staircase with two handrails was fixed adjacent to two force platforms (Fig. 2). The steps' rise and going were 185 mm and 280 mm, respectively. The second step of the staircase was composed of two parts that were individually mounted on different force platforms. To ensure that no noise was captured by the force transducers during the gait's double support phase, 1-cm gaps were left between the second and its two neighbouring steps. The

**Table 2** Participant demographics

Demographics	CR DD	UC	UCR	Control
<i>n</i>	12	5	7	9
Male/female	11/1	1/4	3/4	3/6
Age, mean $\pm$ std (years)	66.6 $\pm$ 4.0	68.4 $\pm$ 8.8	70.1 $\pm$ 6.2	70 $\pm$ 6.4
Affected side (L/R)	5/7	4/1	5/2	-
Pre-op BMI, mean $\pm$ std (kg/m <sup>2</sup> )	<b>30.0 <math>\pm</math> 2.5*</b>	<b>32.0 <math>\pm</math> 5.9*</b>	<b>30.3 <math>\pm</math> 3.3*</b>	24.0 $\pm$ 3
Post-op BMI, mean $\pm$ std (kg/m <sup>2</sup> )	<b>30.1 <math>\pm</math> 2.5*</b>	<b>31.6 <math>\pm</math> 6.1*</b>	<b>29.6 <math>\pm</math> 3.4*</b>	-
Pre-op OKS, mean $\pm$ std	35 $\pm$ 7	42 $\pm$ 4	36 $\pm$ 5	-
Post-op OKS, mean $\pm$ std	<b>24 <math>\pm</math> 8<sup>†</sup></b>	<b>19 <math>\pm</math> 4<sup>†</sup></b>	<b>22 <math>\pm</math> 8<sup>†</sup></b>	-

\* <sup>†</sup> Statistically significant difference between implant and control groups, and between pre-op and post-op visits, respectively ( $P < .01$ )





**Fig. 2** Staircase with a step height and length of 185 mm and 280 mm, respectively

surfaces of the force plates under the second step were also raised by 370 mm in the acquisition software, to compensate for the further displacement from the piezoelectric sensors. Participants were instructed to initiate the trial at a two-step distance from the construction and climb the staircase with their own preferred manner. The use of handrails was not restricted, and no instructions were given as to which foot ought to initiate ascent or descent. Upon reaching the plateau at the top of the staircase, participants were asked to turn around and descend back to the start position. A minimum of five ascent and five descent trials per participant was recorded.

### Processing

Trials were cropped to discard frames with low marker visibility using the Vicon Nexus software. At least one gait cycle per limb per trial was considered. Gaps of less than five frames in the markers' trajectories were filled with Woltring quintic spline fills, whereas larger gaps were filled with rigid body or pattern fills. Marker and force recordings were filtered using a 4th order Butterworth filter with cutoff frequencies of 6 and 300 Hz, respectively. Gait events were automatically detected from force and marker data, and all trial data were normalised to 100% of the gait cycle using custom MATLAB scripts. Finally, all gait cycles for each lower limb were averaged to a single data set per person and type of activity.

Differences in gait parameters between patient groups, both pre-operatively and post-operatively, and the control group were compared. Functional metrics were categorised into spatiotemporal (walking speed, cadence, stride time, contralateral foot-off and heel-strike instances, and duration of double support), kinematic (peak knee joint movements, knee RoM and knee flexion at heel-strike and toe-off) and kinetic parameters (peak knee external moments and power). All kinematic

measurements of the patient and control participants are with reference to the operated and both limbs, respectively.

Since participants ascended and descended the staircase in their own manner, the adopted movement strategies potentially affected the biomechanical functioning. Whilst the use of handrails could expect to reduce the weight bearing of the pathological leg and affect knee joint moments, knee kinematics have been found to be unaffected in old adults [20]. Therefore, stair navigation performance was assessed with the use of spatiotemporal and kinematic parameters only, whilst related kinetic metrics are reported only as supplementary data. Secondly, given that a step-by-step strategy could influence all concerned metrics, only the stair navigation trials of participants following a step-over-step pattern were considered. Strategy preference for the stair navigation tasks (step-by-step or step-over-step) for all groups and visits is also reported.

One-way ANOVAs with a Bonferroni correction were conducted to compare the patients' performance to the controls' (SPSS, IBM, USA). Repeated-measures ANOVAs were also used to assess differences between implant groups at each operative state. The significance level was not adjusted for comparisons and is reported at  $P < 0.05$ ,  $P < 0.01$  and  $P < 0.001$ . In reporting changes, an improvement (highlighted in green in the tables) was defined as a statistically significant change that led to a parameter's average value being closer to the controls' corresponding measurement, whilst the opposite was true for a deterioration in a metric (in red).

## Results

### Walking

With the exception of the post-operative performance of the UC group, all patients, both pre-operatively and post-operatively, had a lower walking speed (m/s) and cadence (steps/min) compared to the controls (Table 3, spatiotemporal parameters). Furthermore, both the CR DD and UCR groups walked with increased stride times during both patient visits. The only significant improvement in the spatiotemporal performance of the post-operative population was observed at the contralateral foot off instance (% of gait cycle) of the CR DD group (Table 3). Contralateral heel strike events (% of gait cycle) were consistent for all groups and visits, and all patient groups had significantly longer double support times compared to the control group.

Pre-operative peak knee joint angles of all implant groups were statistically comparable to the controls' recordings (Table 3) but becoming statistically worse, post-operatively, compared to controls. The difference in knee flexion during the patients' post-operative assessment, compared to before the TKA, was evident

**Table 3** Knee spatiotemporal and kinematic parameters in walking

Parameters		CR DD		UC		UCR		Control
		Pre-Op	Post-Op	Pre-Op	Post-Op	Pre-Op	Post-Op	
Spatio-temporal	Walking speed (m/s)	1.1 ± 0.2***	1.2 ± 0.2*	1.0 ± 0.2**	1.2 ± 0.2	0.9 ± 0.3***	1.1 ± 0.2**	1.5 ± 0.1
	Cadence (steps/min)	103.5 ± 9.8**	104.8 ± 10.3***	104.4 ± 7.6*	111.2 ± 5.5	101.9 ± 14.9**	106.9 ± 10.4**	123.7 ± 9.0
	Stride time (s)	1.2 ± 0.1**	1.2 ± 0.1**	1.2 ± 0.1	1.1 ± 0.1	1.2 ± 0.2**	1.1 ± 0.1*	1.0 ± 0.1
	Contralateral foot off (%)	14.8 ± 3.2*	12.4 ± 1.4 <sup>†</sup>	16 ± 3.0**	13.6 ± 1.5*	15.4 ± 2.4**	14.4 ± 2.0***	10.8 ± 1.4
	Contralateral heel strike (%)	49.4 ± 0.8	49.9 ± 0.8	49.8 ± 0.8	49.3 ± 0.6	50.4 ± 1.0	49.6 ± 1.7	50.1 ± 0.7
	Double support (s)	0.3 ± 0.1**	0.3 ± 0.1*	0.4 ± 0.1**	0.3 ± 0.0*	0.4 ± 0.1**	0.3 ± 0.1**	0.2 ± 0.0
Joint kinematics	Peak flexion (°)	55.7 ± 6.2	54.3 ± 3.4**	53.1 ± 9.6	48.5 ± 5.7***	52.5 ± 7.4	53.7 ± 6.0*	61.2 ± 3.0
	Peak adduction (°)	11.2 ± 4.6	17.9 ± 8.0* <sup>†</sup>	9.1 ± 8.8	12.2 ± 5.4	10.9 ± 4.1	10.3 ± 9.2*	8.4 ± 4.7
	Peak internal rotation (°)	14.1 ± 6.4	8.3 ± 4.0* <sup>††</sup>	9.6 ± 2.5	11.2 ± 3.3	12.0 ± 6.4	8.5 ± 4.9*	17.3 ± 8.0
	Sagittal RoM (°)	51.6 ± 5.3***	56.8 ± 4.6 <sup>†</sup>	49.7 ± 4.3**	52.1 ± 4.4**	49.3 ± 8.2***	53.2 ± 6.6**	62.6 ± 3.2
	Coronal RoM (°)	10.9 ± 6.8	16.4 ± 7.4	11.4 ± 9.1	12.9 ± 7.1	8.7 ± 3.8	13.9 ± 10.7	9.4 ± 5.3
	Transverse RoM (°)	16.2 ± 5.0	13.2 ± 4.6*	11.8 ± 4.4*	15.3 ± 6.7	15.4 ± 5.8	13.5 ± 5.4	20.6 ± 6.1
	Flexion at heel strike (°)	5.6 ± 6.2	-0.3 ± 2.9 <sup>††</sup>	4.9 ± 6.6	-1.8 ± 3.7 <sup>†</sup>	4.5 ± 5.2	3.5 ± 5.7	3.8 ± 4.7
	Flexion at toe off (°)	37.8 ± 5.4	34.6 ± 4.3	36.4 ± 7.8	29.9 ± 4.5*	37.9 ± 7.1	36.4 ± 7.5	38.6 ± 4.3

Flexion, adduction and internal rotation angles are indicated by positive values, while extension, abduction and external rotation by negative numbers

\* / \*\* / \*\*\* = significance at  $p < .05$ ,  $< .01$  and  $< .001$  between implant and control groups, respectively


<sup>†</sup> / <sup>††</sup> / <sup>†††</sup> = significance at  $p < .05$ ,  $< .01$  and  $< .001$  between pre-op and post-op visits

<sup>1,2,3</sup> = significance at  $p < .05$ ,  $< .01$  and  $< .001$  between implant groups on the same visit


throughout the entirety of the gait cycle (Fig. 3). Furthermore, the post-operative relapse of the peak adduction and internal rotation angles (°) of the CR DD was statistically significant (Table 3). Generally, the patients' coronal and transverse RoM (°) was unaffected by the TKA and comparable to the controls' RoM. In contrast, only patients who received a CR DD implant showed a significant improvement in sagittal RoM (°) that was on a par with the control's assessment (Table 3). Finally, both the CR DD and UC bearings displayed knee extension at heel strike, whilst the UC implant demonstrated a significantly reduced flexion at toe off, post-operatively (Table 3).

Knee flexion moment was typically biphasic in nature, with two flexion peaks at early and late stance, and an intermediate extension peak during mid stance (Fig. 4). Knee adduction/abduction moment also followed a

similar pattern with two marginal adductions (Table 4) and one centralised abduction peak (Fig. 4). Throughout the entirety of the gait cycle, normalised pre-operative peak flexion/extension moments for all participants were similar; however, early stance moments for the CR DD group reduced after the operation (Table 4), whilst the analogous mid-stance extension moments increased for the CR DD and UC groups alike. On the whole, peak adduction/abduction moments, during all concerned phases of the gait cycle and patient visits (Table 4), were comparable to the control values, an exemption being the recorded moments in mid-stance from the CR DD group that were significantly increased post-operatively. Finally, both normalised peak power estimates were significantly reduced compared to controls, for all implant types and visits, except for the peak eccentric power of the UCR group at both visits (Table 4).



**Fig. 3** Average (solid lines) knee flexion  $\pm$  standard deviations (shaded bands) for the pre-operative (in red), post-operative (in blue) and control (max/min in black; no average shown) groups. Toe off occurrences (% gait cycle) are indicated by vertical solid lines



**Fig. 4** Average (solid lines) knee moments in walking and  $\pm 2$  standard deviations (shaded bands) for the pre-operative (in red), post-operative (in blue) and control (black; no average shown) groups. Foot off occurrences (% gait cycle) are indicated by vertical solid lines

### Stair navigation

As opposed to the post-operative measurement, pre-operative walking speed in stair navigation was consistently lower compared to the controls' (Tables 5 and 6). Similarly, pre-operative significant differences in the cadence and stride time of the osteoarthritic patients disappeared post-operatively. The CR DD group demonstrated a delayed contralateral foot off timing during stair descent, significantly different to all other study groups (Table 6) which also persisted post-operatively. As for the contralateral heel strike, the pre-operative performance of the UCR group during stair ascent was also significantly different 1 year after surgery (Table 5). For the UC group during stair ascent and descent, and the CR DD group in descent, the pre-operative double support phase was significantly longer compared to the control group, yet the UC implant group exhibited a post-operative reduction in double support time (Table 6).

The control group demonstrated greater peak flexion angles ( $^{\circ}$ ) during stair ascent (Table 5) than all patient groups in all instances. Furthermore, the peak flexion angle of the patients with a UCR bearing decreased 1 year after surgery. Even though the pre-operative peak flexion angles of the UC and UCR groups were equivalent to the control group in stair descent (Table 6), their function declined post-operatively. The adduction angles of the CR DD group significantly increased post-operatively for both stair ascent and descent (Tables 5 and 6), whilst there were no evident differences in peak knee internal rotation angles between groups.

Knee sagittal RoM was higher in stair navigation than during walking. For both stair tasks, the sagittal RoM of the control group was greater than the patients, but only the UCR implant group exhibited a significant post-operative loss of sagittal RoM during stair ascent (Table 5). The CR DD group showed a large and statistically significant post-

**Table 4** Knee kinetic parameters in walking

Parameters		CR DD		UC		UCR		Control
		Pre-Op	Post-Op	Pre-Op	Post-Op	Pre-Op	Post-Op	
Peak flex/ext moment (Nm/kg)	Early Stance	0.4 $\pm$ 0.3	0.1 $\pm$ 0.1**†	0.2 $\pm$ 0.1	0.0 $\pm$ 0.2**	0.3 $\pm$ 0.2	0.2 $\pm$ 0.2	0.4 $\pm$ 0.2
	Mid Stance	0.0 $\pm$ 0.3	0.2 $\pm$ 0.1†	-0.1 $\pm$ 0.2	0.4 $\pm$ 0.2†	0.0 $\pm$ 0.2	0.1 $\pm$ 0.2	0.3 $\pm$ 0.2
	Late Stance	0.3 $\pm$ 0.2	0.2 $\pm$ 0.0***	0.2 $\pm$ 0.1	0.2 $\pm$ 0.1	0.2 $\pm$ 0.1	0.2 $\pm$ 0.0**	0.3 $\pm$ 0.0
Peak add/abd moment (Nm/kg)	Early Stance	0.5 $\pm$ 0.2	0.5 $\pm$ 0.2	0.5 $\pm$ 0.3	0.5 $\pm$ 0.1	0.6 $\pm$ 0.1	0.5 $\pm$ 0.2	0.7 $\pm$ 0.2
	Mid Stance	0.3 $\pm$ 0.2*	0.2 $\pm$ 0.1†	0.3 $\pm$ 0.3	0.2 $\pm$ 0.1	0.4 $\pm$ 0.1*	0.3 $\pm$ 0.2	0.1 $\pm$ 0.1
	Late Stance	0.4 $\pm$ 0.2	0.3 $\pm$ 0.1	0.4 $\pm$ 0.3	0.3 $\pm$ 0.1	0.5 $\pm$ 0.1	0.3 $\pm$ 0.2	0.3 $\pm$ 0.1
Peak concentric power (W/kg)		0.2 $\pm$ 0.1***	0.2 $\pm$ 0.1***	0.1 $\pm$ 0.1***	0.2 $\pm$ 0.1**	0.2 $\pm$ 0.2**	0.2 $\pm$ 0.1**	0.7 $\pm$ 0.4
Peak eccentric power (W/kg)		0.9 $\pm$ 0.3***	0.7 $\pm$ 0.2***	0.6 $\pm$ 0.4***	0.8 $\pm$ 0.5**	0.9 $\pm$ 0.5	0.8 $\pm$ 0.3	1.8 $\pm$ 0.3

\* / \*\* / \*\*\* = significance at  $p < .05$ ,  $< .01$  and  $< .001$  between implant and control groups, respectively

† / †† / ††† = significance at  $p < .05$ ,  $< .01$  and  $< .001$  between pre-op and post-op visits



**Table 5** Knee spatiotemporal and kinematic parameters in stair ascent

	Parameters	CR DD		UC		UCR		Control
		Pre-Op	Post-Op	Pre-Op	Post-Op	Pre-Op	Post-Op	
Spatio-temporal	Walking speed (m/s)	<b>0.5 ± 0.1*</b>	0.6 ± 0.3	<b>0.4 ± 0.1**</b>	0.5 ± 0.1	<b>0.4 ± 0.2**</b>	0.5 ± 0.2	0.6 ± 0.0
	Cadence (steps/min)	77.2 ± 12.2	87.0 ± 17.4	<b>61.5 ± 15.2**</b>	75.2 ± 19.5	<b>62.3 ± 18.1***</b>	71.1 ± 18.9	93.2 ± 6.0
	Stride time (s)	1.6 ± 0.3	1.4 ± 0.4	<b>2.1 ± 0.6*</b>	1.7 ± 0.6	<b>2.1 ± 0.7**</b>	1.7 ± 0.9	1.3 ± 0.1
	Contralateral foot off (%)	16.6 ± 4.4	15.0 ± 3.5	18.5 ± 5.4	14.5 ± 2.6	15.0 ± 4.5	15.2 ± 3.7	12.6 ± 1.3
	Contralateral heel strike (%)	47.7 ± 2.3	49.5 ± 5.0	49.9 ± 1.7	48.4 ± 2.1	<b>52.6 ± 5.5<sup>1,†</sup></b>	<b>45.2 ± 9.3<sup>†</sup></b>	48.0 ± 1.2
	Double support (s)	0.5 ± 0.1	0.4 ± 0.1	<b>0.8 ± 0.5<sup>3</sup></b>	0.5 ± 0.3	0.5 ± 0.1	0.4 ± 0.1	0.3 ± 0.1
Joint kinematics	Peak flexion (°)	<b>85.2 ± 8.6***</b>	<b>84.5 ± 5.0***</b>	<b>88.5 ± 3.7*</b>	<b>82.2 ± 4.3***</b>	<b>89.1 ± 3.5*</b>	<b>76.0 ± 13.6***,††</b>	100.6 ± 4.5
	Peak adduction (°)	20.9 ± 9.0	<b>31.8 ± 9.8<sup>†</sup></b>	14.7 ± 4.6	23.1 ± 10.5	24.4 ± 10.7	21.2 ± 20.6	16.5 ± 8.9
	Peak internal rotation (°)	16.6 ± 7.5	12.5 ± 7.0	12.9 ± 3.7	13.3 ± 3.4	15.6 ± 5.5	11.7 ± 6.0	22.2 ± 11.5
	Sagittal RoM (°)	<b>73.3 ± 11.5*</b>	<b>77.9 ± 5.6*</b>	79.5 ± 8.4	81.6 ± 9.3	83.2 ± 6.6	<b>69.1 ± 11.5***,†</b>	88.2 ± 7.2
	Coronal RoM (°)	17.9 ± 8.7	<b>28.7 ± 8.3<sup>†</sup></b>	16.1 ± 5.3	23.7 ± 10.3	20.2 ± 5.9	24.1 ± 18.9	18.9 ± 4.8
	Transverse RoM (°)	13.3 ± 4.5	12.4 ± 4.7	10.9 ± 3.9	11.4 ± 3.0	15.0 ± 5.4	12.3 ± 8.3	16.7 ± 6.0
	Flexion at heel strike (°)	54.7 ± 8.7	<b>49.3 ± 6.8**</b>	56.8 ± 6.6	<b>48.8 ± 6**</b>	53.8 ± 7.7	<b>47.1 ± 11.3**</b>	63.4 ± 4.6
	Flexion at toe off (°)	15.0 ± 9.5	11.3 ± 3.8	11.4 ± 5.3	<b>3.1 ± 8.7**</b>	7.5 ± 5.7	10.7 ± 7.8	16.7 ± 6.2

Flexion, adduction and internal rotation angles are indicated by positive values, while extension, abduction and external rotation by negative numbers

\* / \*\* / \*\*\* = significance at  $p < .05$ ,  $< .01$  and  $< .001$  between implant and control groups, respectively

† / †† / ††† = significance at  $p < .05$ ,  $< .01$  and  $< .001$  between pre-op and post-op visits

<sup>1,2,3</sup> = significance at  $p < .05$ ,  $< .01$  and  $< .001$  between implant groups on the same visit.

operative increase in the frontal RoM in both ascent and descent (Tables 5 and 6). Finally, except for the UCR group during stair descent, all TKA designs resulted in a straighter leg at heel strike post-operatively compared to the control group.

With regard to the type of movement strategies that were adopted to complete the stair navigation assessment, all control participants paced with a step-over-step strategy, whilst 22% and 38% of the pre-operative recordings showed patients walking step-by-step during stair ascent and descent, respectively (Table 7). TKA surgery also shifted the patients' movement preferences towards the behaviour of the controls during stair ascent (Pearson Chi-square,  $P=0.016$ ) but not during descent ( $P=0.282$ ), in 3% and 12% of the post-operative trials was the step-by-step strategy used whilst ascending and descending, respectively.

## Discussion

This study compared the functional performance of healthy controls and patients who received one of three randomly assigned B. Braun Columbus® knee implants: a low congruency fixed (CR DD), a high congruency fixed (UC), or a high congruency mobile (UCR) bearing. This work is the first in the literature to assess the UCR Columbus mobile bearing in vivo and compare the performance of different Columbus knee prostheses in a series of activities of daily living using motion capture. Spatiotemporal, kinematic and kinetic parameters were compared between all patient groups and controls, between pre-operative and post-operative assessments (intra-implant changes), and between implant groups of the same operative state (inter-implant changes). Participant demographics indicated that controls were similar in age to all patient groups but with a significantly lower BMI (Table 2).

**Table 6** Knee spatiotemporal and kinematic parameters in stair descent

	Parameters	CR DD		UC		UCR		Control
		Pre-Op	Post-Op	Pre-Op	Post-Op	Pre-Op	Post-Op	
Spatio-temporal	Walking speed (m/s)	<b>0.5 ± 0.1***</b>	<b>0.5 ± 0.1*</b>	<b>0.4 ± 0.0**</b>	0.5 ± 0.1	<b>0.5 ± 0.1*</b>	0.5 ± 0.1	0.7 ± 0.1
	Cadence (steps/min)	<b>80.8 ± 12.9***</b>	97.0 ± 24.7	<b>72.2 ± 7.0**</b>	87.5 ± 19.4	<b>78.5 ± 16.8**</b>	77.3 ± 13.4	105.5 ± 5.6
	Stride time (s)	<b>1.5 ± 0.3**</b>	1.3 ± 0.5	<b>1.7 ± 0.2*</b>	1.3 ± 0.5	<b>1.6 ± 0.4*</b>	1.3 ± 0.4	1.1 ± 0.1
	Contralateral foot off (%)	<b>15.1 ± 2.7<sup>1,2</sup></b>	<b>14.8 ± 2.7*</b>	10.9 ± 0.0	11.0 ± 3.3	12.6 ± 2.2	12.8 ± 2.5	10.7 ± 1.3
	Contralateral heel strike (%)	49.1 ± 2.5	51.7 ± 2.9	47 ± 8.9	48.2 ± 2.6	49.0 ± 0.6	46.2 ± 7.7	50.7 ± 2.2
	Double support (s)	<b>0.4 ± 0.1*</b>	<b>0.4 ± 0.1*</b>	<b>0.6 ± 0.0*</b>	<b>0.3 ± 0.1<sup>†</sup></b>	0.4 ± 0.1	0.4 ± 0.0	0.2 ± 0.0
Joint kinematics	Peak flexion (°)	<b>83.6 ± 7.1**</b>	<b>81.7 ± 7.3**</b>	87.8 ± 5.2	<b>80.6 ± 3.2*</b>	82.9 ± 6.7	<b>75.9 ± 12.8***</b>	95.7 ± 6.9
	Peak adduction (°)	17.2 ± 6.8	<b>28.8 ± 11.1*,†</b>	7.6 ± 3.0	22.4 ± 8.5	22.7 ± 14.7	21.3 ± 17.8	12.9 ± 6.1
	Peak internal rotation (°)	16.6 ± 6.8	11.4 ± 6.1	14.3 ± 4.6	13.4 ± 3.7	14.9 ± 6.0	11.7 ± 6.7	21.5 ± 11.0
	Sagittal RoM (°)	<b>77.4 ± 9.2*</b>	<b>79.3 ± 6.1*</b>	86.0 ± 5.6	80.9 ± 6.5	81.4 ± 4.0	<b>73.9 ± 11.3**</b>	89.2 ± 7.2
	Coronal RoM (°)	13.8 ± 8.8	<b>25.0 ± 9.5<sup>†</sup></b>	8.6 ± 3.4	23.2 ± 11.6	14.8 ± 12.4	22.1 ± 17.5	13.6 ± 4.7
	Transverse RoM (°)	17.2 ± 4.3	13.4 ± 4.4	15.1 ± 10.7	12.4 ± 2.4	16.4 ± 1.5	13.9 ± 2.9	18.9 ± 8.3
	Flexion at heel strike (°)	7.2 ± 8.1	<b>3.9 ± 2.8*</b>	1.8 ± 0.3	<b>0.6 ± 2.9**</b>	3.8 ± 7.1	3.5 ± 7.5	10.1 ± 4.1
	Flexion at toe off (°)	82.3 ± 7.4	79.2 ± 6.6	87.1 ± 5.4	78.6 ± 2.7	81.3 ± 7.7	74.7 ± 13.3	86.9 ± 6.1

Flexion, adduction and internal rotation angles are indicated by positive values, while extension, abduction and external rotation by negative numbers

\* / \*\* / \*\*\* = significance at  $p < .05$ ,  $< .01$  and  $< .001$  between implant and control groups, respectively

† / †† / ††† = significance at  $p < .05$ ,  $< .01$  and  $< .001$  between pre-op and post-op visits

<sup>1,2,3</sup> = significance at  $p < .05$ ,  $< .01$  and  $< .001$  between implant groups on the same visit

### Differences between patient and control groups

Statistically significant differences were primarily observed between patient and control groups, rather than amongst implant designs or patient visits. On the whole, patients walked at a slower speed and cadence, and with greater periods of double support. Such differences suggest less confidence in the lower limbs and a reluctance to remain in single support, possibly due to joint pain and discomfort. Although the magnitude of these inequalities was lessened after surgery and the OKS were improved post-operatively, patients' mobility and performance 1 year after surgery had not yet reached the controls' level.

The controls' gait generally featured greater peak knee flexion and rotation angles and lower adduction angles compared to the patient movements across all activities and visits (Tables 3, 5, and 6). Accordingly, controls consistently demonstrated greater sagittal and transverse RoM during all recordings, whilst the coronal RoM was lower than the patient participants in most comparisons. As for the knee flexion at heel strike and toe off, control volunteers generally exhibited higher flexion angles, particularly compared to the patients' post-operative performance. Low knee flexion angles, predominantly during heel strike events in stair navigation, may indicate loss of function and lack of confidence in the operated joint, since straight rather than flexed knee joints may offer more stability in gait initiation. Further, increased knee flexion angles at toe off indicate greater ground clearance prior to swing, thus diminishing the likelihood of tripping and falling.

Knee flexion/extension moments for level walking showed no variability between controls and pre-operative patient participants. Peak adduction moments in early and late stance were also similar across all groups and visits. In contrast, pre-operative peak abduction moments during mid stance were higher for patients compared to the controls; generally, these peak moments were reduced after surgery, whilst the transition of the CR DD group was also statistically significant (Table 4). With the sole exception of the UCR group, peak power magnitudes in walking were significantly lower during all patient visits as compared to the controls.

### Intra-implant changes

In general, patients' functional performance improved following TKA: there were thirty-five statistically

significant differences in the spatiotemporal parameters between pre-operative patients and controls, in contrast to fourteen post-operatively. However, measurements such as walking speed and double support duration never attained statistical equivalence with the control group, suggesting that mobility impairments persisted after surgery. Yet, it is worth mentioning that functional improvements can be made beyond a year after TKA. On three occasions, spatiotemporal parameters were significantly improved 1 year after surgery: the point of contralateral foot off during walking was significantly earlier for the CR DD group, and so was the contralateral heel strike of the UCR group in stair ascent; finally, the period of double support was significantly shorter for the UC implant group in stair descent (Tables 3, 5, and 6, in green). Since both the contralateral heel strike and foot off determine the period of double support, these advancements in the performance of all implant groups imply a degree of confidence in the operated limb to remain longer in single stance support.

In terms of joint kinematics, post-operative patient recordings displayed a generic increase in knee sagittal and coronal RoM, yet the post-operative peak knee flexion and rotation angles were broadly reduced from their pre-operative countervales, implying that the post-operative improved sagittal RoM is due to higher knee extension rather than flexion. Moreover, abnormally high knee adduction angles and coronal RoM (e.g., Tables 5 and 6) are not entirely unexpected: in a pre-operative significantly varus knee, the lateral collateral ligaments may have less strain when loaded in comparison to when the knee is restored to neutral, and the collateral ligaments may be more lax postoperatively. Post-surgery patients displayed a straighter leg at heel strike and toe off, whilst in two instances (CR DD and UC group in walking) the post-operative loss of function for these two parameters was statistically significant.

The post-operative reduced knee flexion angles during heel strike and toe off may have shifted the knee's flexion axis closer to vertical component of the ground reaction force, thus reducing the knee's flexion/extension moment during early and late stance. This was evident in the patients' peak flexion moments in early stance that were generally reduced post-operatively (Table 4), possibly due to the reduced flexion angles at heel strike (Table 3). The UCR group did not show any significant shift in its performance after surgery, whilst

**Table 7** Step-by-step strategy preference for all groups and visits

	CR DD		UC		UCR		All patients		Control
	Pre-Op	Post-Op	Pre-Op	Post-Op	Pre-Op	Post-Op	Pre-Op	Post-Op	
Stair ascent	8%	0%	24%	0%	43%	9%	22%	3%	0%
Stair descent	17%	12%	60%	0%	57%	21%	38%	12%	0%

in contrast the other two patient groups exhibited changes in the moment peaks during early and mid-stance (Table 4). Peak adduction/abduction moments were largely unaffected by the TKA, and only the CR DD witnessed an improvement in the magnitude of the peak abduction moments in mid stance. Finally, patients' post-operative peak power values were in all cases comparable to their pre-operative assessment but consistently lower than the control group.

All healthy controls adopted the step-over-step stair navigation strategy, whilst fewer patients preferred the step-by-step movement strategy after TKA surgery while ascending (Table 7,  $P=0.282$ ), signifying that their post-operative functional improvement was also reflected in their movement behaviour. TKA and osteoarthritis patients' movement behaviour were also previously assessed for other activities of daily living by the same authors [21, 22], similarly indicating that patient participants favour different movement strategies compared to asymptomatic controls.

#### Inter-implant changes

Statistically significant differences between implant design groups were only observed pre-operatively (Tables 5 and 6): the contralateral heel strike in stair ascent for the UCR group, the double support duration in stair ascent for the UC group, and the contralateral foot off in stair descent for the CR DD group. None of these differences persisted 1 year after surgery.

When comparing the total number of parameters that significantly differed between pre-operative and post-operative visits (Tables 3, 4, 5 and 6, key: †), the CR DD implant showed an improvement in four parameters (green-shaded values) and a deterioration in eight (red-shaded values). The UC group had two improved metrics (peak flexion/extension moment in mid stance during walking and double support time during stair descent) and one impaired (flexion at heel strike during walking). Whilst patients with a UCR bearing showed a functional improvement in one parameter (contralateral heel strike % during stair ascent), they also had an adverse change in two (peak flexion and sagittal RoM in stair ascent). In view of the above gross metrics, the UC implant showed an overall functional improvement, whilst the CR DD and UCR bearings were accompanied with an overall deterioration in the patients' function.

The sagittal RoM of the low congruent-fixed knee bearing (CR DD) showed a statistically significant post-operative increase (Table 3, walking trials). Additionally, the same implant group exhibited significantly greater post-operative coronal RoM in both stair ascent and descent, to an extent that it was regarded as a functional disadvantage (Tables 5 and 6, in red). This was anticipated since, in low congruency bearings, the sagittal

femoral radius is decreased in the high end of the flexion range, thus improving the knee's RoM. By contrast, patients with UCR implants did not exhibit an improved tibiofemoral axial rotation during either activity of daily living, despite the design's enhanced transverse rotational freedom of the polyethylene insert on the tibial plateau. Finally, the RoM of the high congruent fixed bearing (UC) on all axes was neither significantly increased nor decreased, for all recorded activities.

#### Study limitations

A limitation of this study emerges from the sample size of the participating population which was further reduced by the high percentage of withdrawals and no-shows, also leading to uneven sample sizes between groups. Even if the study is likely underpowered, a sample size of 24 patients is reasonable for a motion capture study that follows TKA patients for a year and is comparable to previous research on this topic [18, 23, 24]. Additionally, the significant difference in the BMI of the patient and control participants may have significantly influenced the kinematic and kinetic recordings in this work (as described by [25]). Finally, in order to gauge the functional performance of the investigated implant designs, a number of spatiotemporal, kinematic, and kinetic parameters were employed. Even though all studied parameters are relevant to human gait and may be used to distinct natural from pathological movements, they do not carry equal weight in the evaluation of an implant's post-operative functional performance. Additionally, several statistically significant changes amongst patient groups may be regarded as negligible during activities of daily living (e.g., Table 5, increased contralateral heel strike timing of the UCR group) and are potentially clinically insignificant. Therefore, the interpretation of this study's findings is also subject to the clinical significance of each metric.

#### Conclusion

Generally, TKA surgery largely improved the function of all patient groups. Nevertheless, metrics such as the walking speed, double support time and peak power magnitudes never became statistically equivalent to the controls' performance. In terms of the net number of positive post-operative changes, the UC implant provided better post-operative function compared to the other two implants in this study, followed by the UCR and CR DD bearings. Nevertheless, the CR DD group showed the most prominent post-operative improvement during the walking assessment in the most commonly reported functional metric, the sagittal RoM. The enhanced operation range of the CR DD design was attributed in its mechanical properties as a low congruent bearing. Finally, this study is the first in the literature to

assess the Columbus UCR ultra-congruent system with a rotating platform. It was hypothesised that the UCR design would permit more natural movements compared to the other two implants when carrying out everyday activities. Yet, both the UC and UCR groups showed comparable functional performance and improvement after TKA surgery and a similarly restricted post-operative sagittal RoM. Patients with a UCR implant did not exhibit an improvement in their tibiofemoral axial rotation despite the bearing's mobile design.

## Supplementary Information

The online version contains supplementary material available at <https://doi.org/10.1186/s13018-021-02311-x>.

**Additional file 1:** Table 1 Knee kinetic parameters in stair ascent.

**Additional file 2:** Table 2 Knee kinetic parameters in stair descent.

## Acknowledgements

The authors gratefully acknowledge the Golden Jubilee National Hospital, Clydebank, Scotland, for the patient recruitment and financial support.

## Authors' contributions

Dimitrios-Sokratis Komaris: conceptualization, methodology, software, formal analysis, investigation, writing—original draft, visualisation. Cheral Govind: conceptualization, methodology, investigation, software, formal analysis, data curation, writing—review and editing. Andrew James Murphy: conceptualization, methodology, resources, writing—review and editing, supervision. Jon Clarke: conceptualization, methodology, resources, writing—review and editing, funding acquisition. Alistair Ewen: conceptualization, methodology, resources, writing—review and editing. Hollie Leonard: conceptualization, methodology, resources, writing—review and editing. Philip Riches: conceptualization, methodology, resources, writing—review and editing, supervision. The authors read and approved the final manuscript.

## Funding

None.

## Availability of data and materials

Not applicable.

## Ethics approval and consent to participate

Ethical approvals were granted by the Strathclyde University ethics committee and the West of Scotland research ethics committee 5. All participants gave written informed consent.

## Consent for publication

Not applicable. All authors consent for the publications of the manuscript in the *Journal of Orthopaedic Surgery and Research*.

## Competing interests

The authors declare no competing interest.

## Author details

<sup>1</sup>Tyndall National Institute, University College Cork, Lee Maltings Complex Dyke Parade, Cork T12 R5CP, Ireland. <sup>2</sup>Department of Biomedical Engineering, University of Strathclyde, Glasgow, Scotland. <sup>3</sup>School of Medicine, University of Dundee, Dundee, Scotland. <sup>4</sup>Orthopaedic Department, Golden Jubilee National Hospital, Clydebank, Scotland.

Received: 8 February 2021 Accepted: 17 February 2021

Published online: 06 March 2021

## References

- Carr BC, Goswami T. Knee implants – review of models and biomechanics. *Materials & Design*. 2009;30(2):398–413.
- Kim Y-H, Kim D-Y, Kim J-S. Simultaneous mobile- and fixed-bearing total knee replacement in the same patients: a prospective comparison of mid-term outcomes using a similar design of prosthesis. *J Bone Joint Surg British Volume*. 2007;89(7):904–10.
- Victor J, Banks S, Bellemans J. Kinematics of posterior cruciate ligament-retaining and -substituting total knee arthroplasty: a prospective randomised outcome study. *J Bone Joint Surg British volume*. 2005;87(5):646–55.
- Simpson DJ, Gray H, D'Lima D, Murray D, Gill H. The effect of bearing congruency, thickness and alignment on the stresses in unicompartmental knee replacements. *Clin Biomechanics*. 2008;23(9):1148–57.
- Ezzet KA, Garcia R, Barrack RL. Effect of component fixation method on osteolysis in total knee arthroplasty. *Clin Orthop Relat Res*. 1995;321:86–91.
- Feller JA, Bartlett RJ, Lang DM. Patellar resurfacing versus retention in total knee arthroplasty. *The Journal of bone and joint surgery British volume*. 1996;78(2):226–8.
- Lampe F, Sufi-Siavach A, Bohlen KE, Hille E, Dries SPM. One year after navigated total knee replacement, no clinically relevant difference found between fixed bearing and mobile bearing knee replacement in a double-blind randomized controlled trial. *Open Orthop J*. 2011;5:201–8.
- Goebel D, Schultz W. The Columbus Knee System: 4-year results of a new deep flexion design compared to the NexGen full flex implant. *Arthritis*. 2012;2012:213817.
- Jung W-H, Jeong J-H, Ha Y-C, Lee Y-K, Koo K-H. High early failure rate of the Columbus posterior stabilized high-flexion knee prosthesis. *Clin Orthop Relat Res*. 2012;470(5):1472–81.
- Hakki S, Saleh KJ, Potty AG, Bilotta V, Oliveira D. Columbus navigated TKA system: clinical and radiological results at a minimum of 5 years with survivorship analysis. *Orthopedics*. 2013;36(3):e308–18.
- Marques CJ, Daniel S, Sufi-Siavach A, Lampe F. No differences in clinical outcomes between fixed- and mobile-bearing computer-assisted total knee arthroplasties and no correlations between navigation data and clinical scores. *Knee Surgery, Sports Traumatology, Arthroscopy*. 2015;23(6):1660–8.
- Luzo MVM, Ambra LFM, Debieux P, Franciozi CE, Costi RR, Petrilli MdT, et al. Total knee arthroplasty with computer-assisted navigation: an analysis of 200 cases. *Rev Bras Ortop*. 2014;49(2):149–53.
- Kim SH, Lim J-W, Ko Y-B, Song M-G, Lee H-J. Comparison of ultra-congruent mobile- and fixed-bearing navigation-assisted total knee arthroplasty with minimum 5-year follow-up. *Knee Surgery, Sports Traumatology, Arthroscopy*. 2016;24(11):3466–73.
- Lutzner J, Beyer F, Dixel J, Fritzsche H, Lutzner C, Kirschner S. No difference in range of motion between ultracongruent and posterior stabilized design in total knee arthroplasty: a randomized controlled trial. *Knee Surg Sports Traumatol Arthroscopy*. 2017;25(11):3515–21.
- Yoon JR, Yang JH. Satisfactory short-term results of navigation-assisted gap-balancing total knee arthroplasty using ultracongruent insert. *J Arthroplasty*. 2018;33(3):723–8.
- Yoon J-R, Yang J-H. Unsatisfactory clinical outcomes of second-generation mobile bearing floating platform total knee arthroplasty: comparing outcomes with fixed bearing after five years minimum. *Int Orthopaedics*. 2018;42(11):2583–9.
- Fuchs A, Häussermann P, Hömig D, Ochs BG, Müller CA, Helwig P, et al. 5-Jahres-Follow-up von 210 Columbus-Knietotalendoprothesen. *Der Orthopäde*. 2018;47(10):859–66.
- Urwin SG, Kader DF, Caplan N, Gibson ASC, Stewart S. Gait analysis of fixed bearing and mobile bearing total knee prostheses during walking: do mobile bearings offer functional advantages? *The Knee*. 2014;21(2):391–5.
- Jenny JY, Diesinger Y. The Oxford Knee Score: compared performance before and after knee replacement. *Orthopaedics Traumatol Surg Res*. 2012;98(4):409–12.
- Reeves ND, Spanjaard M, Mohagheghi AA, Baltzopoulos V, Maganaris CN. Influence of light handrail use on the biomechanics of stair negotiation in old age. *Gait & Posture*. 2008;28(2):327–36.

21. Komaris D-S, Govind C, Murphy A, Ewen A, Riches P. Identification of movement strategies during the sit-to-walk movement in patients with knee osteoarthritis. *J Appl Biomechanics*. 2018;34(2):96–103.
22. Komaris D-S, Govind C, Clarke J, Ewen A, Jeldi A, Murphy A, et al. Identifying car ingress movement strategies before and after total knee replacement. *International Biomechanics*. 2020;7(1):9–18.
23. Kramers-de Quervain IA, Stüssi E, Müller R, Drobny T, Munzinger U, Gschwend N. Quantitative gait analysis after bilateral total knee arthroplasty with two different systems within each subject. *J Arthroplasty*. 1997;12(2): 168–79.
24. Sosio C, Gatti R, Corti M, Locatelli E, Fraschini G. Motion analysis after total knee arthroplasty. *Sport Sciences for Health*. 2008;4(1):1–6.
25. Stephen PM, Walter HE, Thomas ED, Timothy M, Margaret KJ, Mary LOT, et al. Obesity: effects on gait in an osteoarthritic population. *J Appl Biomechanics*. 1996;12(2):161–72.

## Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

**Ready to submit your research? Choose BMC and benefit from:**

- fast, convenient online submission
- thorough peer review by experienced researchers in your field
- rapid publication on acceptance
- support for research data, including large and complex data types
- gold Open Access which fosters wider collaboration and increased citations
- maximum visibility for your research: over 100M website views per year

**At BMC, research is always in progress.**

Learn more [biomedcentral.com/submissions](https://biomedcentral.com/submissions)

